

DE 3 - 30664

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

SYSTEME CEI D'ACCEPTATION MUTUELLE DE CERTIFICATS D'ESSAIS DES EQUIPEMENTS ELECTRIQUES (IECEE) METHODE OC

CB TEST CERTIFICATE CERTIFICAT D'ESSAI OC

Product

Electronic measuring equipment

Name and address of the applicant

Omicron electronics GmbH

Oberes Ried 1

6833 Klaus, AUSTRIA

Name and address of the manufacturer

Omicron electronics GmbH

Oberes Ried 1, 6833 Klaus, AUSTRIA

Name and address of the factory

Omicron electronics GmbH

Oberes Ried 1, 6833 Klaus, AUSTRIA

Ratings and principal characteristics

ARCO 400

CMC430

Rated voltage: Rated 100-240V

100-240V

frequency:

50/60Hz

50/60Hz

Rated power:

10-8A

1000W

Protection

class:

1

1

Trade mark (if any)

Omicron

Model/type Ref.

ARCO 400 CMC430

Additional information (if necessary)

Certificate DE 3 - 30457 issued 2017-05-23 is replaced by this

version due to technical changes

A sample of the product was tested and found

to be in conformity with

IEC 61010-1:2010 IEC 61010-2-030:2010

as shown in the Test Report Ref. No. which forms part of this certificate

028-713128450-000

This CB Test Certificate is issued by the National Certification Body Ce Certificat d'essai OC est établi par l'Organisme **National de Certification**

Date.

2018-03-26

CB 18 03 17955 108

Es

(lean



Erich Thurner

TÜV SÜD Product Service GmbH · Certification Body · Ridlerstrasse 65 · D-80339 München

Product Service